

SUPERVISOR TEST CONDUCTED ON  
DMT 100166 ON 9/8/2011 @ 10:42:36  
USING A D.100 g/2102 WEST BATH SIMULATOR  
YIELDED THE FOLLOWING RESULTS:

0.6919

0.0932

0.940

942

0.940

0.942

945

0.946

951

1951

REPORTED  
AVERAGE  
0.893

CALCULATED

AVERAGE TO

4 516. Dig-

8941

INITIAL RESULTS IN THIS SERIES ARE  
LOW MOST LIKELY DUE TO CONNECTIVE  
TUBING (UNHEATED) COUPLING THE SIMULATOR  
TO THE FITTING ON THE INSTRUMENT. HAD  
THE CONNECTIVE TUBING NOT BEEN USED  
(WHICH IS THE NORMAL MODE OF USE IN VT)  
THE AVERAGE WOULD HAVE MOST LIKELY  
BEEN HIGHER. CAL AS RECEIVED WAS  
0.950187 FROM CALIBRATION ON 10/19/2010.  
RECALIBRATION AT NPAS YIELDED A CAL  
OF 0.996355

David Padonacki 9/9/11